

HC5D-B

(Cat No.1516)

HEMATOLOGY CONTROLS

CONTROL

LOT E5228

EXP. 2025-12-16

Instrument : SYSMEX XN-Series / XN-9000 *										
Parameter / Paramètre		LOW			NORMAL			HIGH		
		Mean Cibles	±	Limit Limites	Mean Cibles	±	Limit Limites	Mean Cibles	±	Limit Limites
WBC/GB	10 ³ /µL & 10 ⁹ /L	3,60	±	1,20	7,15	±	1,80	19,75	±	3,00
WBC-D/GB-D	10 ³ /µL & 10 ⁹ /L	3,70	±	1,20	7,80	±	1,80	19,95	±	3,00
RBC/GR	10 ⁶ /µL & 10 ¹² /L	2,19	±	0,15	4,57	±	0,24	5,81	±	0,30
Hgb	g/dL	5,9	±	0,4	13,4	±	0,6	18,2	±	0,7
	g/L	59	±	4	134	±	6	182	±	7
	mmol/L	3,7	±	0,2	8,3	±	0,2	11,3	±	0,2
Hct	%	17,5	±	2,0	39,5	±	2,5	53,2	±	3,0
	L/L	0,175	±	0,020	0,395	±	0,025	0,532	±	0,030
MCV/VGM	fL	80,0	±	5,0	86,5	±	5,0	91,5	±	5,0
MCH/TCMH	pg	26,9	±	2,5	29,3	±	2,5	31,3	±	2,5
	fmol	1,7	±	0,2	1,8	±	0,2	1,9	±	0,2
MCHC/CCMH	g/dL	33,7	±	3,0	33,9	±	3,0	34,2	±	3,0
	g/L	337	±	30	339	±	30	342	±	30
	mmol/L	21,1	±	1,9	21,0	±	1,9	21,2	±	1,9
RDW/IDR-SD	fL	50,0	±	10,0	49,0	±	10,0	50,0	±	10,0
RDW/IDR-CV	%	18,5	±	5,0	15,0	±	5,0	15,0	±	5,0
Plt	10 ³ /µL & 10 ⁹ /L	48	±	20	185	±	40	395	±	65
MPV/VPM	fL	9,6	±	3,0	8,2	±	3,0	8,0	±	3,0
NEUT#	10 ³ /µL & 10 ⁹ /L	2,07	±	0,36	3,65	±	0,72	10,96	±	1,97
LYMPH#	10 ³ /µL & 10 ⁹ /L	0,59	±	0,32	1,82	±	0,64	4,54	±	1,78
MONO#	10 ³ /µL & 10 ⁹ /L	0,29	±	0,29	0,46	±	0,46	0,79	±	0,79
EO#	10 ³ /µL & 10 ⁹ /L	0,49	±	0,33	0,89	±	0,64	2,57	±	1,78
BASO#	10 ³ /µL & 10 ⁹ /L	0,16	±	0,16	0,32	±	0,32	0,89	±	0,89
NEUT%	%	57,5	±	10,0	51,0	±	10,0	55,5	±	10,0
LYMPH%	%	16,5	±	9,0	25,5	±	9,0	23,0	±	9,0
MONO%	%	8,0	±	8,0	6,5	±	6,5	4,0	±	4,0
EO%	%	13,5	±	9,0	12,5	±	9,0	13,0	±	9,0
BASO%	%	4,5	±	4,5	4,5	±	4,5	4,5	±	4,5
NRBC#	10 ³ /µL & 10 ⁹ /L	0,18	±	0,18	0,07	±	0,07	1,38	±	1,38
NRBC%	%	5,0	±	5,0	1,0	±	1,0	7,0	±	7,0
RET% (1)	%	0,50	±	0,50	0,80	±	0,80	1,00	±	1,00
RET# (1)	10 ⁶ /µL & 10 ¹² /L	0,0110	±	0,0110	0,0366	±	0,0366	0,0581	±	0,0581
IRF (1)	%	20,0	±	20,0	20,0	±	20,0	20,0	±	20,0
RET-He (1)	pg	30,0	±	20,0	30,0	±	20,0	30,0	±	20,0

Plt-O	10 ³ /µL & 10 ⁹ /L	NA	NA	NA
Plt-F	10 ³ /µL & 10 ⁹ /L	44 ± 38	232 ± 65	402 ± 85
Pct	%	0,05 ± 0,05	0,15 ± 0,07	0,32 ± 0,11
	mL/L	0,5 ± 0,5	1,5 ± 0,7	3,2 ± 1,1
PDW/IDP	fL	9,0 ± 5,0	7,5 ± 5,0	7,5 ± 5,0
P-LCR/RGC	%	23,0 ± 20,0	20,0 ± 20,0	20,0 ± 20,0
IPF	%	19,5 ± 15,0	19,5 ± 15,0	19,5 ± 15,0
IG#	10 ³ /µL & 10 ⁹ /L	0,52 ± 0,18	0,89 ± 0,36	2,67 ± 0,99
IG%	%	14,5 ± 5,0	12,5 ± 5,0	13,5 ± 5,0
RBC/GR-O	10 ⁶ /µL & 10 ¹² /L	2,18 ± 0,30	4,50 ± 0,50	5,70 ± 0,60
LFR (1)	%	80,0 ± 20,0	80,0 ± 20,0	80,0 ± 20,0
MRF (1)	%	14,0 ± 14,0	14,0 ± 14,0	14,0 ± 14,0
HRF (1)	%	6,0 ± 6,0	6,0 ± 6,0	6,0 ± 6,0